



Certificate of Traceability

Manufacturer Certification

This certifies that **Tosoh Corporation**, using **G8 1.6 min Variant Analysis Mode** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#7**

The system evaluated was:

Instrument: HLC-723G8	Calibrator Lot: ZS0002	Column Lot: B
Reagent Lot: B8-104R, B8-204R, B8-303R, HW-18R	Calibrator Assigned Values: 6.03%, 10.98%	

Date of Certification: June 1, 2011

Certification Expires: June 1, 2012

David Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor