



Certificate of Traceability

Manufacturer Certification

This certifies that **Tosoh Corporation**, using **G7 2.2 min Variant Analysis Mode** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#7**

The system evaluated was:

Instrument: HLC-723G7	Calibrator Lot: ZS0002	Column Lot: H
Reagent Lot: H7-101R, H7-201R, 7-301R, HW-18R	Calibrator Assigned Values: 6.03%, 10.98%	

Date of Certification: June 1, 2011

Certification Expires: June 1, 2012

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor