



Certificate of Traceability

Manufacturer Certification

This certifies that **Tosoh Corporation**, using **ST AIA-Pack HbA1c (AIA-360)** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#7**

The system evaluated was:

Instrument: AIA-360	Calibrator Lot: #5751-2
Reagent Lot: J860011, J570068, I480023, J715785	Calibrator Assigned Values: 0.0, 3.3, 6.6, 9.8, 13.0, 17.3%

Date of Certification: June 1, 2010

Certification Expires: June 1, 2011

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor