



Certificate of Traceability

Manufacturer Certification

This certifies that Tosoh Corporation, using 1.6 min Variant Analysis Mode Ver.5.24F G8 has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial Reference** method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: G8	Calibrator Lot: ZS0004	Column Lot: S
Reagent Lot: S8-105G, S8-209G, S8-308G, HW-34G	Calibrator Assigned Values: 5.68%, 10.61%	

Date of Certification: June 1, 2022

Certification Expires: June 1, 2023

Handwritten signature of David Sarkis in black ink.

NGSP Steering Committee Chair

Handwritten signature of Randie R Little PhD. in black ink.

NGSP Network Coordinator

Handwritten signature of Sharon Anulf in black ink.

SRL director/ supervisor